

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0183 Mueller Hinton Agar
Batch Number: 00003630
Expiry Date: 05.02.2023
Date Received: 07.02.2020
Date Tested: 07.02.2020
Sample Condition: Satisfactory
Sample Number: 372335

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP0972 and PP0963 with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Haemophilus influenzae NCTC 12975/ATCC 49766	90	Opaque/grey colonies	Opaque/grey colonies
Streptococcus pneumoniae NCTC 12977/ATCC 49619	80	Grey/green colonies with alpha-haemolysis	Grey/green colonies with alpha-haemolysis

Physical	Result	Specifications
pH	7.4	7.3 +/- 0.1
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Cherry Red88-1 - 88-3

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0183 Mueller Hinton Agar
Batch Number: 00003630
Expiry Date: 05.02.2023
Date Received: 07.02.2020
Date Tested: 07.02.2020
Sample Condition: Satisfactory
Sample Number: 372335

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

All of the results on this certificate of analysis relate only of the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd