

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0183 Mueller Hinton Agar  
**Batch Number:** 00004378  
**Expiry Date:** 14.04.2024  
**Date Received:** 24.05.2021  
**Date Tested:** 24.05.2021  
**Sample Condition:** Satisfactory  
**Sample Number:** 403204

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Product sample prepared and tested as PP0972 with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Haemophilus influenzae NCTC 12975/ATCC 49766	98	Opaque/grey colonies	Opaque/grey colonies
Streptococcus pneumoniae NCTC 12977/ATCC 49619	109	Grey/green colonies with alpha-haemolysis	Grey/green colonies with alpha-haemolysis

Physical	Result	Specifications
pH	7.3	7.3 +/- 0.1
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Cherry Red88-1 - 88-3

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0183 Mueller Hinton Agar  
**Batch Number:** 00004378  
**Expiry Date:** 14.04.2024  
**Date Received:** 24.05.2021  
**Date Tested:** 24.05.2021  
**Sample Condition:** Satisfactory  
**Sample Number:** 403204

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

All of the results on this certificate of analysis relate only of the samples submitted.



**Douglas Cameron**  
Technical Manager, E&O Diagnostics Ltd