

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: KM0004 CLED Single Indicator Agar
Batch Number: 00008090
Expiry Date: 17/01/2027
Date Received: 09/02/2024
Date Tested: 09/02/2024
Date of Issue: 15/05/2024
Sample Condition: Satisfactory

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP0080 CLED Single Indicator Agar

For Solid Media the RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / $\geq 50\%$ for a selective media.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
E.coli NCTC 12241	Growth	Yellow colonies	Yellow colonies
E.faecalis NCTC 12697	Growth	Pale yellow colonies	Pale yellow colonies
K.pneumoniae NCTC 9633	Growth	Yellow colonies	Yellow colonies
P.aeruginosa NCTC 12903	Growth	Blue colonies	Blue colonies
P.mirabilis NCTC 10975	Growth	Blue colonies	Blue colonies
P.vulgaris ATCC 8427	Growth	Blue colonies	Blue colonies
S.aureus NCTC 12981	Growth	Yellow colonies	Yellow colonies

Physical	Result	Specification	Test Method
Colour	Conforms	Green/blue. 2241 C, 2247 C, 2401 C	ED/SOP/009 by visual observation. Range measured using Pantone guide.

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: KM0004 CLED Single Indicator Agar

Batch Number: 00008090

Expiry Date: 17/01/2027

Date Received: 09/02/2024

Date Tested: 09/02/2024

Date of Issue: 15/05/2024

Sample Condition: Satisfactory

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

pH	7.2	7.3 ± 0.2	ED/SOP/003 measurement by pH meter.
----	-----	-----------	-------------------------------------

All of the results on this certificate of analysis relate only to the samples submitted. Test specifications are based on ISO 11133:2014/Amd.2:2020 and internal product specifications

Douglas Cameron
Technical Manager, E&O Laboratories Ltd