

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

**Sample:** BM0320 - Rappaport-Vass. Broth  
**Batch Number:** 04084359  
**Expiry Date:** 2020-10-09  
**Date Received:** 2020-04-10  
**Date Tested:** 2020-04-10  
**Sample Condition:** Satisfactory  
**Sample Number:** 376168

### Accredited Test Method: See table below

Productivity	Result	Colonial Appearance	Colonial Appearance Specification
Salmonella typhimurium NCTC 12023	Growth	Translucent colonies with black centre	Translucent colonies with black centre

For solid media the RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media /  $\geq 50\%$  for selective media. For liquid media growth is determined from a subculture of the test broth inoculated with 10 - 100 cfu of the target organism.

### Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor $\geq 2$ .

Selectivity	Test Result	Specification
Escherichia coli NCTC 12241	Inhibited	Inhibited
Enterococcus faecalis NCTC 12697	Inhibited	Inhibited
Pseudomonas aeruginosa NCTC 12903	Inhibited	Inhibited

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999. The inspection level is $\geq 0.4\%$ of the batch and the reject level $\leq 7$ units depending on batch size.			
pH	5.3	5.2 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	DE224-1C - DE238-1C Dark Blue	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Diagnostics Ltd

**Sample:** BM0320 - Rappaport-Vass.Broth  
**Batch Number:** 04084359  
**Expiry Date:** 2020-10-09  
**Date Received:** 2020-04-10  
**Date Tested:** 2020-04-10  
**Sample Condition:** Satisfactory  
**Sample Number:** 376168

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Fill Quantity	Conforms	6ml +/- 0.3	ED/SOP/054 by gravimetric determination
---------------	----------	-------------	---

Accredited Test Method for Solid Media	Accredited Test Method for Liquid Media	Accredited Test Method for Diluents
ED/SOP/008 Quantitative evaluation using spread inoculum technique	ED/SOP/008 Semi quantitative inoculation with growth assessment from subculture	ED/SOP/008 Semi quantitative evaluation of viability maintenance after a holding time of 45 minutes

All of the results on this certificate of analysis relate only to the samples submitted.



Douglas Cameron  
Technical Manager, E&O Diagnostics Ltd