

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** BM1673 - Virus Transport Medium RT  
**Batch Number:** 04084788  
**Expiry Date:** 2021-05-07  
**Date Received:** 2020-05-08  
**Date Tested:** 2020-05-08  
**Sample Condition:** Satisfactory  
**Sample Number:** 377612

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

| Physical   | Result   | Specifications                      | Accredited Test Method  |
|--|----------|-------------------------------------|---|
| Sterility  | Conforms | Within acceptable limits            | ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C |
| Sterility sampling is performed in accordance with ISO 2859-1:1999. The inspection level is $\geq 0.4\%$ of the batch and the reject level $\leq 7$ units depending on batch size. |          |                                     |   |
| pH   | 7.4      | 7.2 +/- 0.2                         | ED/SOP/003 measurement by pH meter  |
| Colour   | Conforms | DE87-6 - DE90-6 Liquid / Pale Peach | ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide            |
| Fill Quantity  | Conforms | 3ml +/- 0.3                         | ED/SOP/054 by gravimetric determination   |

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Diagnostics Ltd

**Sample:** BM1673 - Virus Transport Medium RT  
**Batch Number:** 04084788  
**Expiry Date:** 2021-05-07  
**Date Received:** 2020-05-08  
**Date Tested:** 2020-05-08  
**Sample Condition:** Satisfactory  
**Sample Number:** 377612

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

### Accredited Test Method for Solid Media

ED/SOP/008 Quantitative evaluation using spread inoculum technique

### Accredited Test Method for Liquid Media

ED/SOP/008 Semi quantitative inoculation with growth assessment from subculture

### Accredited Test Method for Diluents

ED/SOP/008 Semi quantitative evaluation of viability maintenance after a holding time of 45 minutes

All of the results on this certificate of analysis relate only to the samples submitted.



Douglas Cameron  
Technical Manager, E&O Diagnostics Ltd